





**Certificate:** 80218075

**Master Contract:** 261694

**Project:** 80218075

**Date Issued:** 2024-10-23

---

3. Environmental Conditions: Extended (Specify): 0°C to 40°C, 20% to 80%RH, non-condensing

**APPLICABLE REQUIREMENTS**

CSA C22.2 No. 61010-1-12, UPD1:2015, UPD2:2016, AMD1:2018, UPD3:2023 - Safety requirements for electrical equipment for measurement, control, and laboratory use — Part 1: General requirements - Third Edition; Update No. 1: July 2015; Update No. 2: April 2016; Update No. 3: June 2023

UL 61010-1 3rd ed (Rev. Jun 6, 2023) - UL Standard for Safety Electrical Equipment For Measurement, Control, and Laboratory Use; Part 1: General Requirements - Third Edition; Reprint with Revisions Through and Including JUNE 6, 2023



**Certificate:** 80218075


**Master Contract:** 261694

**Project:** 80218075

**Date Issued:** 2024-10-23

Notes:

Products certified under Class(es) C363106, C363186 have been certified under CSA's ISO/IEC 17065 accreditation with the Standards Council of Canada (SCC). [www.scc.ca](http://www.scc.ca)





## *Supplement to Certificate of Compliance*

**Certificate:** 80218075

**Master Contract:** 261694

*The products listed, including the latest revision described below,  
are eligible to be marked in accordance with the referenced Certificate.*

### **Product Certification History**

---

<b>Project</b>	<b>Date</b>	<b>Description</b>
80218075	2024-10-23	cCSAus certification of Programmable DC Power Supply