



Certificate of Compliance

Certificate: 80178207

Master Contract: 261694

Project: 80178207

Date Issued: 2025-01-07

Issued To: ITECH Electronic Co Ltd
108 Xi Shan Qiao Nan Lu
Nanjing, Jiangsu, 210039
China

Attention: Lily Han

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US' for Canada and US or with adjacent indicator 'US' for US only or without either indicator for Canada only.

Issued by: *Magnus Chen*
Magnus Chen



PRODUCTS

CLASS - C363106 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT Electrical Measurement and Test Equipment

CLASS - C363186 - ELECTRICAL MEASUREMENT AND TEST EQUIPMENT Certified to US Standards

AC/DC Power Supply

Model(s)	Input Voltage (V)	Frequency (Hz)	Current (A max)	Output Voltage (V)	Current (A)	Power (kVA)
IT7803-350-30U	3~, 200-220 or 380-480	50/60	20	0-350VAC L-N (single phase only) 0-499VDC	30	3



Certificate: 80178207
Project: 80178207

Master Contract: 261694
Date Issued: 2025-01-07

IT7803W-350-30U	3~, 200-220 or 380-480	50/60	20	0-350VAC L-N (single phase only) 0-499VDC	30	3
IT7803J-350-30U	1~, 220	50/60	20	0-350VAC L-N (single phase only) 0-499VDC	30	3
IT7805-350-30U	3~, 200-220 or 380-480	50/60	20	0-350VAC L-N (single phase only) 0-499VDC	30	5
IT7806-350-90	3~, 200-220 or 380-480	50/60	25	0-350VAC L-N 0-499VDC	90	6
IT7808-350-90	3~, 200-220 or 380-480	50/60	30	0-350VAC L-N 0-499VDC	90	8
IT7809-350-90	3~, 200-220 or 380-480	50/60	34	0-350VAC L-N 0-499VDC	90	9
IT7812-350-90	3~, 200-220 or 380-480	50/60	29	0-350VAC L-N 0-499VDC	90	12
IT7815-350-90	3~, 200-220 or 380-480	50/60	34	0-350VAC L-N 0-499VDC	90	15
IT7815-350-90-Master	3~, 200-220 or 380-480	50/60	34	0-350VAC L-N 0-499VDC	90	15
IT7815-350-90-Slave	3~, 200-220 or 380-480	50/60	34	0-350VAC L-N 0-499VDC	90	15

Notes:

1. The above model is permanently connected, Equipment Class I, Pollution Degree 2, Installation Category II.
2. Mode of operation: Continuous
3. Environmental Conditions: Operating temperature 0 to 50°C, Operating altitude up to 2000m max, Operating humidity 20% to 80% RH (non-condensing)

Regenerative Grid Simulator

Model(s)	Input Voltage (V)	Frequency (Hz)	Current (A max)	Output Voltage (VAC)	Current (A)	Power (kVA)
IT7905-350-30U	3~, 200-220 or 380-480	50/60	20	0-350VAC L-N (single phase only)	30	5



Certificate: 80178207
Project: 80178207

Master Contract: 261694
Date Issued: 2025-01-07

				0-499VDC		
IT7906-350-90	3~, 200-220 or 380-480	50/60	25	0-350VAC L-N 0-499VDC	90	6
IT7908-350-90	3~, 200-220 or 380-480	50/60	30	0-350VAC L-N 0-499VDC	90	8
IT7909-350-90	3~, 200-220 or 380-480	50/60	34	0-350VAC L-N 0-499VDC	90	9
IT7912-350-90	3~, 200-220 or 380-480	50/60	29	0-350VAC L-N 0-499VDC	90	12
IT7915-350-90	3~, 200-220 or 380-480	50/60	34	0-350VAC L-N 0-499VDC	90	15
IT7915-350-90-Master	3~, 200-220 or 380-480	50/60	34	0-350VAC L-N 0-499VDC	90	15
IT7915-350-90-Slave	3~, 200-220 or 380-480	50/60	34	0-350VAC L-N 0-499VDC	90	15
IT7905P-350-30U	3~, 200-220 or 380-480	50/60	20	0-350VAC L-N (single phase only) 0-499VDC	30	5
IT7906P-350-90	3~, 200-220 or 380-480	50/60	25	0-350VAC L-N 0-499VDC	90	6
IT7908P-350-90	3~, 200-220 or 380-480	50/60	30	0-350VAC L-N 0-499VDC	90	8
IT7909P-350-90	3~, 200-220 or 380-480	50/60	34	0-350VAC L-N 0-499VDC	90	9
IT7912P-350-90	3~, 200-220 or 380-480	50/60	29	0-350VAC L-N 0-499VDC	90	12
IT7915P-350-90	3~, 200-220 or 380-480	50/60	34	0-350VAC L-N 0-499VDC	90	15
IT7915P-350-90-Master	3~, 200-220 or 380-480	50/60	34	0-350VAC L-N 0-499VDC	90	15
IT7915P-350-90-Slave	3~, 200-220 or 380-480	50/60	34	0-350VAC L-N 0-499VDC	90	15

Notes:

1. The above model is permanently connected, Equipment Class I, Pollution Degree 2, Installation Category II.
2. Mode of operation: Continuous
3. Environmental Conditions: Operating temperature 0 to 50°C, Operating altitude up to 2000m max, Operating humidity 20% to 80% RH (non-condensing)



Certificate: 80178207
Project: 80178207

Master Contract: 261694
Date Issued: 2025-01-07

Regenerative AC/DC Electronic Load

Model(s)	Output Voltage (V)	Frequency (Hz)	Current (A max)	Input Voltage (V)	Current (A)	Power (kVA)
IT8203-350-30U	3~, 200-220 or 380-480	50/60	14	0-350VAC L-N (single phase only) 0-499VDC	30	3
IT8205-350-30U	3~, 200-220 or 380-480	50/60	15	0-350VAC L-N (single phase only) 0-499VDC	30	5
IT8206-350-90	3~, 200-220 or 380-480	50/60	15	0-350VAC L-N 0-499VDC	90	6
IT8208-350-90	3~, 200-220 or 380-480	50/60	21	0-350VAC L-N 0-499VDC	90	8
IT8209-350-90	3~, 200-220 or 380-480	50/60	23	0-350VAC L-N 0-499VDC	90	9
IT8212-350-90	3~, 200-220 or 380-480	50/60	21	0-350VAC L-N 0-499VDC	90	12
IT8215-350-90	3~, 200-220 or 380-480	50/60	26	0-350VAC L-N 0-499VDC	90	15
IT8215-350-90-Master	3~, 200-220 or 380-480	50/60	26	0-350VAC L-N 0-499VDC	90	15
IT8215-350-90-Slave	3~, 200-220 or 380-480	50/60	26	0-350VAC L-N 0-499VDC	90	15

Notes:

1. The above model is permanently connected, Equipment Class I, Pollution Degree 2, Installation Category II.
2. Mode of operation: Continuous
3. Environmental Conditions: Operating temperature 0 to 50°C, Operating altitude up to 2000m max, Operating humidity 20% to 80% RH (non-condensing)



Certificate: 80178207
Project: 80178207

Master Contract: 261694
Date Issued: 2025-01-07


APPLICABLE REQUIREMENTS

CAN/CSA-C22.2 No. 61010-1-12
UPD1: 2015, UPD2: 2016, AMD1:
2018
UL Std. No. 61010-1 (3rd Edition)
(2012), AMD1:2018

- Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements
- Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements

Notes:

Products certified under Class(es) C363106 have been certified under CSA's ISO/IEC 17065 accreditation with the Standards Council of Canada (SCC). www.scc.ca





Supplement to Certificate of Compliance

Certificate: 80178207

Master Contract: 261694

*The products listed, including the latest revision described below,
are eligible to be marked in accordance with the referenced Certificate.*

Product Certification History

Project	Date	Description
80178207	2025-01-07	cCSAus certification of High Power Programmable AC/DC Power Supply, Regenerative Grid Simulator, Regenerative AC/DC Electronic Load.